


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-9A-N-LK*

S.N.: *8637*

Date: *AUG 27 1998*



- a) Output signal amplitude:
0 TO -200 mA
- b) Pulse width:
0.4 TO 4.0 NS
- c) Rise time:
≤ 200 PS
- d) Fall time:
≤ 200 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
350 mA.

DI MAN-FRAME
OUT TO 40 db
10 VOLTS / DIV

Dot PRF-51 MIE OUT
(TO 20db)
200 mA / DIV
1 NS / DIV
PRF = 300 kHz