


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MP-M-19-C-PW-05*  
 S.N.: *8644*  
 Date: *NOV 9 1998*



- a) Output signal amplitude: *0 to ±5 volts*
- b) Pulse width: *0.2 to 4.0 ns*
- c) Rise time: *≤ 60 ps*
- d) Fall time: *≤ 135 ps*
- e) PRF: *0 to 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*

*50-60 Hz*

*40 dB ATTEN*  
*5 volts/div*  
*TOP Port, 1 ns/div*  
*MID Port, 100 ps/div (RISE TIME)*  
*BOT Port, 500 ps/div*  
*PRF = 200 kHz*