


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-3A-P5-P-05*

S.N.: *8649*

Date: *NOV 24 1998*



- a) Output signal amplitude: *0 to +300 VOLTS*
- b) Pulse width: *150 ns*
- c) Rise time: *5 to 100 ns*
- d) Fall time: *≤ 600 ps*
- e) PRF: *≤ 3 ns*
- f) Jitter, stability: *OK*

70 dB ATTEN
320 VOLTS/DIV
TOP 10 ns/DIV
BOT 1 ns/DIV
(RISE TIME)
PRF = 3 kHz

g) Prime power: *170/240 V*
50-60 Hz
[Signature]