


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV14-MPT-EGGIA

S.N.: 8660

Date: OCT 29 1998



- a) Output signal amplitude:
 ≥ 20 VOLTS (TO 50m)
- b) Pulse width:
 (FWHM) 350 ps
- c) Rise time:
 NA
- d) Fall time:
 NA
- e) PRF: 0 TO 100 MHz
- f) Jitter, stability:
 OK
- g) Prime power:

+15 VDC
100 mA (MAX)

40 db ATTEN
10 VOLTS / DIV
PRF = 300 kHz.