


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVN-1-C-PN-M-OT*

S.N.: *2663*

Date: *DEC 2 1998*



- a) Output signal amplitude: *0 to ± 5 Volts*
- b) Pulse width: *(50ns) 0.2 to 1.0 ns*
- c) Rise time: *≤ 100 ps*
- d) Fall time: *≤ 135 ps*
- e) PRF: *50 to 250 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*

*50-60 MHz*  
*[Signature]*

*40 dB ATTEN*  
*5 Volts/DIV*  
*PRF ≈ 100 MHz*  
*TDR 500 ps/DIV*  
*BOI 5 ns/DIV*