


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MVH-5-MV1-P*

S.N.: *8670*

Date: *DEC 10 1998*



- a) Output signal amplitude: *0 TO +40 V*
- b) Pulse width: *~350 ps (FWHM)*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 V DC*

40 db ATTEN

10 VOLTS/DIV

PRF = 300 kHz

+15 V DC

200 mA (max)

A