


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-C-PN-OT-EX-EW-ED-EO*

S.N.: *8676*

Date: *DEC 17 1998*



- a) Output signal amplitude:
0 TO ±5 VOLTS
- b) Pulse width:
0.2 TO 5.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
5 VOLTS/DIV
PRF = 12 MHz

120/240V
50-60Hz
[Signature]