PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-3AH-C-P-M-EA

S.N.: 8691

Date: FEB 19 1999



a) Output signal amplitude: ✓
0 TO + 450 VOLTS

b) Pulse width: (@ 50Ω)
5 TO 100 NS

c) Rise time:
≤ 1 NS ✓

d) Fall time:
≤ 4 NS ✓

e) PRF:
0 TO 5 KHZ ✓

f) Jitter, stability:
OK ✓

g) Prime power:
120 / 240 ✓ VOLTS
50 - 60 HZ

70 db ATTEN

320 VOLTS /DIV

PRF = 2 KHZ