


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVR-61-C-PN-URA


S.N.: 8696

Date: DEC 7 1998



- a) Output signal amplitude: 0 TO  $\pm 500$  VOLTS
- b) Pulse width:  $R_L \geq 100K$ , 1  $\mu$ S TO 100 MS  
(20% MAX DUTY CYCLE)
- c) Rise time:  $\leq 20$  NS
- d) Fall time:  $\leq 20$  NS
- e) PRF: 0 TO 1 KHz  
(20% MAX DUTY CYCLE)
- f) Jitter, stability: OK
- g) Prime power: 120/240V  
50-60 Hz.

(A) Point  $R_L = 100K$ .



(B) Newt,  $R_L = 100K$   
100 V/DIV  
1 MS/DIV.