


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2-B-N-EA*

S.N.: *8699*

Date: *FEB 15 1999*



- a) Output signal amplitude: *070-350V (50ns)*
- b) Pulse width: *5 TO 100 NS*
- c) Rise time: *≤ 2 NS*
- d) Fall time: *≤ 3 NS*
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*

≈ 60 dB ATTEN
≈ 100 V/DIV / DIV
TDR 10 NS / DIV
BDI 2 NS / DIV
(RISE TIME)
PRF = 2 KHz.

50-60V
✓