


PULSE GENERATOR
PERFORMANCE CHECK

Model: ANM-2-C-P

S.N.: 3703

Date: JAN 25 1999



- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width: (50%)
0.2 TO 2.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 25 MHz.
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60Hz

40 db ATTEN

10 VOLTS/DIV

TDP 1 NS/DIV

VBOT 100 ps/DIV

PRF = 10 MHz