


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-1-PW-PS-P-DA-EW*

S.N.: *8716*

Date: *FEB 8 1999*



- a) Output signal amplitude: *0 to +200 V*
- b) Pulse width: *($R_L \geq 50 \Omega$)
50 ns to 5.0 μ s*
- c) Rise time: *≤ 10 ns*
- d) Fall time: *≤ 10 ns*
- e) PRF: *0 to 1 kHz.*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V
50 - 100 Hz*

*$R_L = 50 \Omega$
PRF = 1 kHz*

TOP *1 μ s/DIV*

BOT *100 ns/DIV.*

*120/240 V
50 - 100 Hz
[Signature]*