


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AUP-AU-1413-P-05*

S.N.: *8726*

Date: *JAN 7 1997*



*50 dB ATTEN*  
*32 VOLTS/DIV*  
*PRF ≈ 300 KHz*

- a) Output signal amplitude:  
*0 TO +40 VOLTS*
- b) Pulse width: *(50 ns)*  
*0.4 TO 2.0 NS*
- c) Rise time:  
*≤ 150 ps*
- d) Fall time:  
*≤ 250 ps*
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 V DC*  
*300 mA MAX*

*[Signature]*