


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-6A1-C-P*


S.N.: *8730*

Date: *FEB 2 1999*




- a) Output signal amplitude: *0 to +1.0 Amp*
- b) Pulse width: *50 ns to 5.0 us*
- c) Rise time: *1 ns*
- d) Fall time: *1 ns*
- e) PRF: *0 to 10 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 Hz

*(A) MAINFRMNG OUT TO 50db - 32 V/DIV
TR 50 NS/DIV
MID 2 NS/DIV (RISE TIME)
BOT 2 NS/DIV (FALL TIME)*



50 NS/DIV
10 NS/DIV



(B)