


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MHF-1*
S.N.: *8731*
Date: *DEC 17 1998*



- a) Output signal amplitude: $\geq -10 \text{ VOLTS}$
- b) Pulse width: $\leq 130 \text{ ps}$
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: $100 \pm 5 \text{ MHz}$
- f) Jitter, stability: *OK*
- g) Prime power:

*0.5 WATT (15 VOLTS P-P)
100 MHz CW.*

*40 db ATTN
10 VOLTS/DIV
TRR 2 NS/P10*

BSI 200 ps/P10