


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVIR-2-PS-N*

S.N.: *8739*

Date: *FEB 17 1999*



- a) Output signal amplitude:
0 TO -200 V (50Ω)
- b) Pulse width:
2 TO 10 NS
- c) Rise time:
≤ 1 NS
- d) Fall time:
PW MIN: ≤ 1 NS
PW MAX: ≤ 2 NS
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/200 V
50-60 Hz

60 db atten
~ 100 V/DIV

120/200 V
50-60 Hz
[Signature]