


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AUR-A-1-C-52-P-PLA*

S.N.: *8742*

Date: *FEB 11 1999*



- a) Output signal amplitude: *0 TO +225 V*
- b) Pulse width: *(TO 50ns)*
- c) Rise time: *10 TO 100 ns*
- d) Fall time: *≤ 3 ns*
- e) PRF: *0 TO 50 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*

50-60 kHz
~~*100*~~

60 dB ATTEN
-100 VOLTS DIV.
PRF = 50 kHz.