


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AV-1015-B

S.N.: 8743

Date: December 22, 1998



-50V amplitude, 100 ns PW, 2.5 MHz  
20V/div, 100 ns/div

a) Output Signal Amplitude: 0 to  $\pm 50V$


b) Pulse Width: 20 ns - 10 ms

c) Rise Time: 10 ns

d) Fall Time: 10 ns

e) PRF: 1Hz - 10 MHz

f) Jitter, Stability: OK



+50V amplitude, 10 ms PW, 25 Hz  
20V/div, 2.5 ms/div

g) Prime Power: 120/240V AC, 50-60 Hz,