


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-P*

S.N.: *8744*

Date: *FEB 9 1999*



- a) Output signal amplitude:
0 TO +15V (50%)
- b) Pulse width:
0.2 TO 2.0 ns
- c) Rise time:
5100ps
- d) Fall time:
5135ps
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 VDC
600 mA MAX

40db ATTN
10 VOLT/DIV
PRF ≈ 10 MHz

[Signature]