


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-2-P*

S.N.: *8745*

Date: *FEB 9 1999*



- a) Output signal amplitude:
0 TO +15 V (50 Ω)
- b) Pulse width:
0.2 TO 2.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 VDC

40 Ω 16 MHz

10 VOLT/DIV

PRF ≈ 10 MHz

600 mA MAX

MF