


PULSE GENERATOR  
PERFORMANCE CHECK

1999

Model: *AVL-3AM-C-P-T-M*

S.N.: *8750*

Date: *APR 27 1999*



a) Output signal amplitude:  
*0 TO +450 V (47)*

b) Pulse width:

*(TO 50ns)*

*MIN TR* 5 TO 100 NS

c) Rise time:

*≤ 1 NS*

*VARIOUS TR SETTINGS* d) Fall time:

*≤ 4 NS*

PRF:

*0 TO 5 KHz*

f) Jitter, stability:

*OK*

g) Prime power:

*120/240 V*

*50-60 Hz*

*[Signature]*

*70 dB ATTEN*

*∴ ≈ 320 V/DIV*