


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVG-3A-PS D*

S.N.: *8754*

Date: *JAN 21 1999*



- a) Output signal amplitude: *0 TO + 450 VOLTS*
- b) Pulse width: *(50 ns)*  
*≤ 4 NS (20% RISE)*
- c) Rise time: *≤ 1 NS*  
*(P40)*
- d) Fall time: *≤ 2.5 NS*
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*  
*50-60 Hz*

*60 db ATTN*  
*(50 ns)*

*100 VOLTS/DIV*

*PRF = 10 KHz*

*50-60 Hz*