


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANN-1-C-P*

S.N.: *8765*

Date: *JAN 27 1999*



- a) Output signal amplitude: *0 TO +5 VOLTS*
- b) Pulse width: *(50ns) 0.2 TO 1.0 NS*
- c) Rise time: *≤ 100ps*
- d) Fall time: *≤ 100ps*
- e) PRF: *50 TO 250 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*

40 dB ATTEN
10 VOLTS/DIV
TR 2NS/DIV
BOT 200ps/DIV
PRF = 100 MHz

50-160 VOLTS
[Signature]