


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-4A P*

S.N.: *8766*

Date: *MAR 3 1999*



- a) Output signal amplitude: *0 to +650 V*
- b) Pulse width: *(50ns)*
± 5ns
- c) Rise time:
- d) Fall time:
- e) PRF: *0 to 10KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 VOLT*
750mA MAX

60 db ATTN
100 VOLTS / DIV.
PRF = 5 kHz