


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MN-107B-C-PN-M*

S.N.: *8769*

Date: *JAN 26 1999*



- a) Output signal amplitude:  
*0 to ± 2 AMP*  
*(to 0 to ± 60 V)*
- b) Pulse width:  
*20 to 200 NS*
- c) Rise time:  
*≤ 10 NS*
- d) Fall time:  
*≤ 10 NS*
- e) PRF: *0 to 20 K Hz.*

*PG 7, R<sub>L</sub> = 21 Ω*

f) Jitter, stability: *OK*

g) Prime power:

*120/240 V*  
*5 → 160 Hz*

*TOP MONITOR TO 50 Ω 2V/DIV*  
*BOI LOAD VOLTAGE, 20V/DIV*



*PG 4, R<sub>L</sub> = 21 Ω*