


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-P5-N*

S.N.: *8775*

Date: *FEB 22 1999*



- a) Output signal amplitude:
0 TO -15V
- b) Pulse width: *(50Ω)*
0.2 TO 2.0 NS
- c) Rise time:
≤ 100ps.
- d) Fall time:
≤ 135ps.
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60Hz.

40db ATTN
10 V0VdB/PN
PRF ≈ 10 MHz.