


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMMR-2-P*

S.N.: *8807*

Date: *MMAR 9 1999*



- a) Output signal amplitude:
0 TO +10 V (50Ω)
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 1.0 MHz

- f) Jitter, stability:
OK
- g) Prime power:
*+24 VOLT
300 mA
MAX*

40 dB ATTEN
10 V/DIV
10 NS/DIV
200 ps/DIV (RISE TIME)
200 ps/DIV (FALL TIME)
PRF = 300 kHz

MJ