


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANMP-2-P*

S.N.: *3800*

Date: *MAR 9 1999*



- a) Output signal amplitude:  
*0 to +10 VOLTS*
- b) Pulse width: *(50ns)*
- c) Rise time: *5 to 100 ns*
- d) Fall time: *≤ 100 ps*
- e) PRF: *≤ 135 ps*  
*0 to 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+ 24 VOLTS*  
*300mA MAX*

*40 dB ATTN*  
*10 VOLTS/DIV*  
*TOP 10 ns/DIV*  
*MID 200 ps/DIV (RISE TIME)*  
*BOI 200 ps/DIV (FALL TIME)*  
*PRF ≈ 300 kHz*

*MA*