


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-CN*

S.N.: *8816*

Date: *MAR 8 1999*



- a) Output signal amplitude:
0 TO -10V (50%)
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability:
OK

g) Prime power:
120/240V
50-60 Hz
OK AT 100V ± 1%

40 db ATTEN
10 V/13 DIV
TOP 1NS/DIV
BOT 200ps/DIV

[Signature]