


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MR-3-B-F*

S.N.: *8819*

Date: *APRIL 6 1999*



- a) Output signal amplitude: *0 TO +200 VOLTS*
- b) Pulse width: *TO $R_L \geq 50 \Omega$*
- c) Rise time: *0.1 TO 100 μ S*
(0.5% MAX PULSE CYCLE)
 ≤ 10 NS
- d) Fall time: *≤ 10 NS*
- e) PRF: *0 TO 10 KHz*
(0.5% MAX PULSE CYCLE)
- f) Jitter, stability: *0 IL*
- g) Prime power: *120/240 V*
50-60 Hz

70V 50 V/DIV
10 μ S/DIV
 $R_L = 50 \Omega$
PRF = 50 kHz

BoE 50 V/DIV
100 NS/DIV
 $R_L = 50 \Omega$
PRF = 10 KHz

[Signature]