


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-C-N-1A-0A*

S.N.: *8820*

Date: *APRIL 22 1999*



- a) Output signal amplitude: *0 to -50 V*
- b) Pulse width: *(to 50%)*
2 to 100 ns
- c) Rise time:
≤ 350 ps
- d) Fall time:
2 ns
- e) PRF: *0 to 20 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *100 ± 10*

VOUT.
50 - 60 kHz.

50 dB ATTEN
32 V/283 /DIV
TOP: 10 ns / DIV
BOT: 500 ps / DIV
(RISE TIME):