


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVG-3B-P*

S.N.: *8852*

Date: *APRIL 23 1999*



- a) Output signal amplitude:
+450 V MAX
- b) Pulse width: *(50 ns)*
≤ 2 ns
- c) Rise time:
≤ 0.8 ns
- d) Fall time:
≤ 0.8 ns
- e) PRF:
0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC
350 mA MAX

60 db ATTN
≈ 100 VOLT/DIV
PRF = 10 kHz