


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVG-3B-P*

S.N.: *8854*

Date: *APRIL 23 1999*



- a) Output signal amplitude:  
*+450 V MAX*
- b) Pulse width: *(50 ns)*  
 $\leq 2 \text{ NS}$
- c) Rise time:  
 $\leq 0.8 \text{ NS}$
- d) Fall time:  
 $\leq 0.8 \text{ NS}$
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability:  
*O.K.*
- g) Prime power:

*60 db ATTN*  
*~ 100 VOLT/PW*  
*PRF = 10 KHz*

*+15 VDC*  
*350 mA MAX*