


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMP-27*
S.N.: *8859*
Date: *APRIL 20 1999*



- a) Output signal amplitude:
0 TO +10 V
- b) Pulse width: *(50%)*
5 TO 100 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF: *0 TO 10 MHz*
- f) Jitter, stability:
OK
- g) Prime power:
+24 VDC

40 db ATTN
- 10 V/V/div

TOP: 10 ns/div

RISE: 200 ps/div (RISE TIME)

FALL: 200 ps/div (FALL TIME)

300 mA (MAX)
[Signature]