


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMP-M 1402-PS-DP-05-ET-EN-D*

S.N.: *8867*

Date: *MAY 27 1999*



a) Output signal amplitude:

0 TO +14 V

b) Pulse width:

0 TO 2 NS

c) Rise time:

≤ 300 ps

d) Fall time:

≤ 300 ps

e) PRF:

0 TO 1 MHz

f) Jitter, stability: *OK*

g) Prime power:

*120/240V
50-60 Hz*

*40 db ATTEN
- - 10 VOLTS/DIV*

