


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMP-AV-14V2-05-EA-EW-N*

S.N.: *8868*

Date: *MAY 27 1989*



- a) Output signal amplitude: *0 TO -20 V (TO 50Ω)*
- b) Pulse width: *0.3 TO 2.0 NS*
- c) Rise time: *≤ 100ps*
- d) Fall time: *≤ 200ps*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 V DC*  
*250 mA (MAX)*

*40 db ATTN*

*10 VOLTS/DIV*

*PRF = 300 kHz*