


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-10-2-W-C-P*

S.N.: *8872*

Date: *APRIL 19 1969*



- a) Output signal amplitude: *0 TO +10 V*
- b) Pulse width: *(50 ns)*
- c) Rise time: *20 TO 500 NS*
≤ 150 ps
- d) Fall time: *≤ 200 ps*
- e) PRF: *0 TO 50 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*

50-60 Hz
[Signature]

40 dB ATTEN
10 VOUT/DIV
TOP: 50 NS/DIV
MID: 200 ps/DIV (RISE TIME)
BOT: 200 ps/DIV (FALL TIME)