


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-EI-P-6D1

S.N.: 3890

Date: JUNE 3 '99



- a) Output signal amplitude:
0 TO +20 VOLTS (50 Ω)
- b) Pulse width:
50 NS TO 500 NS
- c) Rise time:
≤ 300 ps
- d) Fall time:
≤ 300 ps
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK

- g) Prime power:
+24 VDC
300 mA (MAX)
- [Signature]*

40 dB ATTEN
10 VOLTS/DIV
100 50 NS/DIV
150T 500 PS/DIV (RISE TIME)
PRF = 30 KHz