


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR ~~ET~~-P-6D1

S.N.: 8891

Date: JUNE 3 1999



- a) Output signal amplitude:
0 to +20V (50 μ V)
- b) Pulse width:
50ns to 500ns
- c) Rise time:
 ≤ 300 ps
- d) Fall time:
 ≤ 300 ps
- e) PRF:
0 to 50 KHz.
- f) Jitter, stability:
OK
- g) Prime power:

+24VDC
300mA (MAX)

40db ATTEN
10 V/DIV
50 ns/DIV
500ps/DIV (RISE TIME)

PRF = 30 KHz.