


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MP-AV-1P*

S.N.: *3907*

Date: *JUN 13 1999*



- a) Output signal amplitude:
0 TO +10 V
- b) Pulse width: *(50-50)*
0.2 TO 4.0 NS
- c) Rise time:
≤ 100ps
- d) Fall time:
≤ 135ps
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
-10 V0 283/PDN
PRF = 300 kHz

+15 VDC
300 mA (MAX)