


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANP-AN-1-C-P-M-OT-401*

S.N.: *3909*

Date: *JULY 15 1999*



40 db ATTN
5 VOLTS/DIV
BMF = 300 kHz
TIP 50 ps/DIV
BOI 2 ns/DIV

- a) Output signal amplitude:
0 TO +5V (50Ω)
- b) Pulse width:
0.25 TO 4.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 125 ps
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 MHz
[Signature]