


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-MP-P-URA

S.N.: 8912

Date: JUN 7 1999



a) Output signal amplitude:

$\geq +40$ VOLTS ✓

b) Pulse width:

1 TO 100 NS ✓

c) Rise time:

≤ 350 PS ✓

d) Fall time:

≤ 2 NS ✓

e) PRF:

MIN PW: 0 TO 1 MHz ✓

MAX PW: 0 TO 20 kHz ✓

f) Jitter, stability:

OK ✓

g) Prime power: ✓

+15 VDC ✓

100 mA (MAX) ✓

50 dB ATTEN

$\therefore 32$ VOLTS/DIV

PRF = 30 kHz