


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-MP-P-URA

S.N.: 89B

Date: JUN 7 7 1999



- a) Output signal amplitude:
 $\geq +40$ VOLTS ✓
- b) Pulse width:
1 TO 100 NS ✓
- c) Rise time:
 ≤ 350 PS ✓
- d) Fall time:
 ≤ 2 NS ✓
- e) PRF:
MIN PW: 0 TO 1 MHz ✓
MAX PW: 0 TO 20 kHz ✓
- f) Jitter, stability:
OK ✓
- g) Prime power: ✓
+15 VDC ✓
100 mA (MAX) ✓

50 dB ATTEN

$\therefore 32$ VOLTS/DIV

PRF = 30 kHz