


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-HUB-GE03*

S.N.: *8916*

Date: *JULY 19 1999*



- a) Output signal amplitude: ✓
± 20 V RMS (± 50 V)
- b) Pulse width: ✓
≤ 350 ns
- c) Rise time: ✓
≤ 100 ns
- d) Fall time: ✓
≤ 200 ns
- e) PRF: ✓
0 to 10 MHz
- f) Jitter, stability: ✓
OK
- g) Prime power: ✓

40 db ATTN
10 V RMS/DIV
PRF = 300 kHz

- a) *+15 V, 35 mA MAX*
- b) *-15 V, -35 mA MAX*
- c) *+135 V, 4 mA MAX*