


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-H23-GE03*

S.N.: *8917*

Date: *JUL 19 1999*



- a) Output signal amplitude: \checkmark
 $\pm 20 \text{ Vrms (}\approx 50 \Omega)$
- b) Pulse width: \checkmark
 $\leq 350 \text{ ps}$
- c) Rise time: \checkmark
 $\leq 100 \text{ ps}$
- d) Fall time: \checkmark
 $\leq 200 \text{ ps}$
- e) PRF: \checkmark
 $0 \text{ to } 10 \text{ MHz}$
- f) Jitter, stability: \checkmark
OK

- g) Prime power: \checkmark
 - a) $+15 \text{ V, } 35 \text{ mA MAX}$
 - b) $-15 \text{ V, } -35 \text{ mA MAX}$
 - c) $+135 \text{ V, } 4 \text{ mA MAX}$

40 db ATTN
-10 VOLTS/PIV
PRF = 300 kHz

[Signature]