


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-HU3-GE03*

S.N.: *3919*

Date: *AUG 19 1999*



- a) Output signal amplitude: $\pm 20 \text{ V}$ ($\pm 2.5 \text{ V}$)
- b) Pulse width: $\leq 350 \text{ ps}$
- c) Rise time: $\leq 100 \text{ ps}$
- d) Fall time: $\leq 200 \text{ ps}$
- e) PRF: 0.90 1.0 MHz
- f) Jitter, stability: *OK*

40 dB ATTEN
10 V/DIV
PRF = 300 kHz

- g) Prime power:
 - a) $+15 \text{ V}$, 35 mA MAX
 - b) -15 V , 35 mA MAX
 - c) $+135 \text{ V}$, 4 mA MAX

[Signature]