


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVX-DD-A1-PS-UVE

S.N.: 8923

Date: June 11, 1999



a) Output Signal Amplitude: TTL, $R_L \geq 50\Omega$

b) Pulse Width: 200 ns

c) Delay Step Size: 27.7, 55.5, 111, or 222ns

d) PRF: 0 - 2 kHz

e) Jitter, Stability: OK

f) Prime Power: 120/240V AC, 50-60 Hz,

Top: Sync Output (buffered input signal)

Bottom: Delayed Output,
Scanning Direction switch set to "Up"
Thumbwheel set to 4
Step Size = 27.7 ns
Lock switch set to "Reset"