


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVG-3-PRI-P*

S.N.: *2929*

Date: *OCT 4 1999*



- a) Output signal amplitude:
+150 to +500 V
- b) Pulse width: *(to 50ns)*
≤ 5 ns (1/2 power)
- c) Rise time:
< 2 ns
- d) Fall time:
< 10 ns
- e) PRF:
0 to 50 kHz
- f) Jitter, stability:
OK
- g) Prime power:

60 db ATTEN
≈ 100 V/DIV / DIV
PRF = 50 kHz

+15 to +100VDC
550mA (max)