


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-C-7

S.N.: 8936

Date: JULY 27 1999



- a) Output signal amplitude: 0 TO +350 V TO 50Ω
- b) Pulse width: 5 TO 100 NS
- c) Rise time: ≤ 2 NS
- d) Fall time: ≤ 3 NS
- e) PRF: 0 TO 5 KHz
- f) Jitter, stability: OK
- g) Prime power: 120/240 V

60 db ATTN

100 VOLT/DIV

PRF = 2 KHz

50-60 Hz