


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AV-151B-C

S.N.: 8948

Date: JUNE 10 1999



- a) Output signal amplitude: 0 TO ± 200V ✓
- b) Pulse width: CR ≥ 50A ✓  
SINE, SQUARE, TRIANGLE
- c) Rise time: +DC OFFSET ✓  
≤ 145 ✓
- d) Fall time: ≤ 145 ✓
- e) PRF: 0 TO 100KHz ✓
- f) Jitter, stability: ✓
- g) Prime power: OK ✓  
120 / 240V ✓  
50 to 60, Hz ✓

1045/DIV  
50V/DIV  
R<sub>L</sub> = 50K

[Signature]